

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091924\
 Data File : VN084045.D
 Acq On : 19 Sep 2024 19:41
 Operator : JC\MD
 Sample : P4003-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE120D3-20240912

Quant Time: Sep 20 01:36:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	144109	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	247031	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	205080	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	78908	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	115507	53.157	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 106.320%			
35) Dibromofluoromethane	8.171	113	81142	51.308	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.620%			
50) Toluene-d8	10.565	98	299932	54.860	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 109.720%			
62) 4-Bromofluorobenzene	12.847	95	102341	45.371	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 90.740%			
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.377	101	25278	15.316	ug/l	98
12) 1,1-Dichloroethene	4.347	96	10979	7.044	ug/l	89
27) cis-1,2-Dichloroethene	7.483	96	5723	2.886	ug/l	93
44) Trichloroethene	9.353	130	1217600	733.890	ug/l	91
55) 1,1,2-Trichloroethane	11.018	97	802	0.490	ug/l	94
64) Tetrachloroethene	11.106	164	9800	7.186	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091924\
Data File : VN084045.D
Acq On : 19 Sep 2024 19:41
Operator : JC\MD
Sample : P4003-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE120D3-20240912

Quant Time: Sep 20 01:36:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

