

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091924\
 Data File : VN084046.D
 Acq On : 19 Sep 2024 20:05
 Operator : JC\MD
 Sample : P4003-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE103D1-20240912

Quant Time: Sep 20 01:37:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	145684	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	256818	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	220489	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	90338	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	117034	53.277	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	106.560%		
35) Dibromofluoromethane	8.171	113	82063	49.913	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.820%		
50) Toluene-d8	10.565	98	312390	54.961	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	109.920%		
62) 4-Bromofluorobenzene	12.847	95	109910	46.870	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	93.740%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.365	101	14150	8.481	ug/l #	20
12) 1,1-Dichloroethene	4.330	96	2106	1.337	ug/l #	78
27) cis-1,2-Dichloroethene	7.500	96	2626	1.310	ug/l #	66
44) Trichloroethene	9.353	130	936902	543.184	ug/l	91
64) Tetrachloroethene	11.100	164	9373	6.392	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091924\
Data File : VN084046.D
Acq On : 19 Sep 2024 20:05
Operator : JC\MD
Sample : P4003-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE103D1-20240912

Quant Time: Sep 20 01:37:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

