

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091924\
 Data File : VN084048.D
 Acq On : 19 Sep 2024 20:53
 Operator : JC\MD
 Sample : P4003-13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE103D3-20240912

Quant Time: Sep 20 01:37:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	146041	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	254060	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	216651	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	86932	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	116216	52.775	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 105.560%			
35) Dibromofluoromethane	8.165	113	82125	50.493	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.980%			
50) Toluene-d8	10.571	98	310346	55.195	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 110.380%			
62) 4-Bromofluorobenzene	12.847	95	108200	46.642	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 93.280%			
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.371	101	3392	2.028	ug/l #	89
44) Trichloroethene	9.353	130	258672	151.597	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091924\
Data File : VN084048.D
Acq On : 19 Sep 2024 20:53
Operator : JC\MD
Sample : P4003-13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE103D3-20240912

Quant Time: Sep 20 01:37:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

