

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091924\
 Data File : VN084049.D
 Acq On : 19 Sep 2024 21:18
 Operator : JC\MD
 Sample : P4003-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE126D2-20240912

Quant Time: Sep 20 01:38:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	141725	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	246345	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	206804	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	81942	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	114320	53.495	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.000%			
35) Dibromofluoromethane	8.165	113	79774	50.584	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.160%			
50) Toluene-d8	10.565	98	296923	54.461	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 108.920%			
62) 4-Bromofluorobenzene	12.853	95	99311	44.151	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 88.300%			
Target Compounds						
44) Trichloroethene	9.353	130	596806	360.718	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091924\
Data File : VN084049.D
Acq On : 19 Sep 2024 21:18
Operator : JC\MD
Sample : P4003-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE126D2-20240912

Quant Time: Sep 20 01:38:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

