

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091924\
 Data File : VN084054.D
 Acq On : 19 Sep 2024 23:18
 Operator : JC\MD
 Sample : P4003-20
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUPE03-20240913

Quant Time: Sep 20 01:40:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	137602	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	239577	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	200028	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	73380	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	109760	52.900	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 105.800%			
35) Dibromofluoromethane	8.165	113	79197	51.636	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.280%			
50) Toluene-d8	10.565	98	284521	53.661	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 107.320%			
62) 4-Bromofluorobenzene	12.847	95	93365	42.680	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 85.360%			
Target Compounds						
44) Trichloroethene	9.353	130	27492	17.086	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091924\
Data File : VN084054.D
Acq On : 19 Sep 2024 23:18
Operator : JC\MD
Sample : P4003-20
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUPE03-20240913

Quant Time: Sep 20 01:40:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

