

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092018\
 Data File : VN051343.D
 Acq On : 20 Sep 2018 16:44
 Operator : MD\SY
 Sample : VN0920WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0920WBL01

Quant Time: Sep 21 06:54:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092018W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 21 03:27:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	746737	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1051440	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	914866	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	356255	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	389860	48.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.34%	
35) Dibromofluoromethane	7.59	113	399033	46.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.92%	
50) Toluene-d8	10.09	98	1449171	44.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.72%	
62) 4-Bromofluorobenzene	12.40	95	408544	38.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	76.88%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092018\
Data File : VN051343.D
Acq On : 20 Sep 2018 16:44
Operator : MD\SY
Sample : VN0920WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0920WBL01

Quant Time: Sep 21 06:54:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092018W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 21 03:27:45 2018
Response via : Initial Calibration

