

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092018\
 Data File : VN051350.D
 Acq On : 20 Sep 2018 19:44
 Operator : MD\SY
 Sample : PB113045ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113045ZHE#01

Quant Time: Sep 21 07:12:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092018W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 21 03:27:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	817162	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1192562	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1019668	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	400228	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	449514	50.75	ug/l	0.00
Spiked Amount				50.000		
Recovery				=	101.50%	
35) Dibromofluoromethane	7.59	113	446182	46.30	ug/l	0.00
Spiked Amount				50.000		
Recovery				=	92.60%	
50) Toluene-d8	10.09	98	1619753	43.72	ug/l	0.00
Spiked Amount				50.000		
Recovery				=	87.44%	
62) 4-Bromofluorobenzene	12.40	95	451994	37.50	ug/l	0.00
Spiked Amount				50.000		
Recovery				=	75.00%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	29066	15.975	ug/l	100
18) Methyl Acetate	4.33	43	8371	1.339	ug/l	90
20) Methylene Chloride	4.55	84	39338	4.321	ug/l	93
43) Isopropyl Acetate	8.17	43	289537	27.376	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092018\
 Data File : VN051350.D
 Acq On : 20 Sep 2018 19:44
 Operator : MD\SY
 Sample : PB113045ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113045ZHE#01

Quant Time: Sep 21 07:12:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092018W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 21 03:27:45 2018
 Response via : Initial Calibration

