

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092018\
 Data File : VN051357.D
 Acq On : 20 Sep 2018 22:36
 Operator : MD\SY
 Sample : J4778-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SA-01-091918

Quant Time: Sep 21 07:24:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092018W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 21 03:27:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	773611	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1164316	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	977727	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	393386	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	440793	52.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.14%	
35) Dibromofluoromethane	7.59	113	431860	45.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.80%	
50) Toluene-d8	10.09	98	1550468	42.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.72%	
62) 4-Bromofluorobenzene	12.40	95	433376	36.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.66%	
Target Compounds						
16) Acetone	3.82	43	49953	29.001	ug/l	99
18) Methyl Acetate	4.33	43	12469	2.107	ug/l #	85
20) Methylene Chloride	4.55	84	348883	40.479	ug/l	95
43) Isopropyl Acetate	8.17	43	285805	27.679	ug/l #	89
95) Naphthalene	15.13	128	30956	2.962	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092018\
Data File : VN051357.D
Acq On : 20 Sep 2018 22:36
Operator : MD\SY
Sample : J4778-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SA-01-091918

Quant Time: Sep 21 07:24:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092018W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 21 03:27:45 2018
Response via : Initial Calibration

