

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092019\
 Data File : VN058244.D
 Acq On : 20 Sep 2019 13:02
 Operator : JC/SP
 Sample : VN0920WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0920WBL02

Quant Time: Sep 21 00:22:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	486917	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	781897	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	650387	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	262239	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	340474	50.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.40%	
35) Dibromofluoromethane	7.57	113	240647	50.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.26%	
50) Toluene-d8	10.09	98	962343	51.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.92%	
62) 4-Bromofluorobenzene	12.41	95	305291	44.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.76%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092019\
Data File : VN058244.D
Acq On : 20 Sep 2019 13:02
Operator : JC/SP
Sample : VN0920WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0920WBL02

Quant Time: Sep 21 00:22:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

