

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092019\
 Data File : VN058249.D
 Acq On : 20 Sep 2019 14:53
 Operator : JC/SP
 Sample : K4974-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE132D2-20190916

Manual Integrations
 APPROVED

MMDadoda
 9/23/2019 3:30:12 PM

Quant Time: Sep 23 04:51:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	499758	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	814549	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	686866	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	278236	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	360040	51.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.46%	
35) Dibromofluoromethane	7.57	113	245718	49.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.24%	
50) Toluene-d8	10.09	98	1018992	52.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.64%	
62) 4-Bromofluorobenzene	12.40	95	321036	44.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.60%	
Target Compounds						
2) Dichlorodifluoromethane	1.83	85	8251	2.010	ug/l	91
9) 1,1,2-Trichlorotrifluoroet	3.74	101	14766	3.655	ug/l	97
27) cis-1,2-Dichloroethene	6.82	96	2487m	0.491	ug/l	
44) Trichloroethene	8.83	130	11413	2.599	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092019\
Data File : VN058249.D
Acq On : 20 Sep 2019 14:53
Operator : JC/SP
Sample : K4974-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE132D2-20190916

Manual Integrations
APPROVED

MMDadoda
9/23/2019 3:30:12 PM

Quant Time: Sep 23 04:51:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

