

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092019\
 Data File : VN058252.D
 Acq On : 20 Sep 2019 15:58
 Operator : JC/SP
 Sample : K4975-06DL 5X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE120D3-20190917DL

Quant Time: Sep 23 05:08:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	477092	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	769394	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	659967	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	267158	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	344810	51.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.78%	
35) Dibromofluoromethane	7.57	113	235925	50.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.88%	
50) Toluene-d8	10.09	98	968185	53.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.26%	
62) 4-Bromofluorobenzene	12.41	95	309241	45.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.36%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.74	101	3820	0.990	ug/l #	78
44) Trichloroethene	8.83	130	149874	36.139	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092019\
Data File : VN058252.D
Acq On : 20 Sep 2019 15:58
Operator : JC/SP
Sample : K4975-06DL 5X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE120D3-20190917DL

Quant Time: Sep 23 05:08:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

