

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092019\
 Data File : VN058255.D
 Acq On : 20 Sep 2019 17:04
 Operator : JC/SP
 Sample : K4975-05DL 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE120D2-20190917DL

Quant Time: Sep 23 05:26:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	487831	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	800581	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	679043	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	269562	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	350702	51.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.24%	
35) Dibromofluoromethane	7.58	113	244591	50.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.52%	
50) Toluene-d8	10.08	98	1004306	52.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.92%	
62) 4-Bromofluorobenzene	12.41	95	315518	44.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.58%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.74	101	7908	2.005	ug/l	95
44) Trichloroethene	8.82	130	327427	75.877	ug/l	98
64) Tetrachloroethene	10.63	164	2409	0.728	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092019\
Data File : VN058255.D
Acq On : 20 Sep 2019 17:04
Operator : JC/SP
Sample : K4975-05DL 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE120D2-20190917DL

Quant Time: Sep 23 05:26:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

