

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092019\
 Data File : VN058258.D
 Acq On : 20 Sep 2019 18:10
 Operator : JC/SP
 Sample : K4974-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB03-20190916

Manual Integrations
 APPROVED

MMDadoda
 9/23/2019 3:30:42 PM

Quant Time: Sep 23 05:45:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	502893	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	808867	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	694361	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	283160	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	360608	51.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.96%	
35) Dibromofluoromethane	7.57	113	249334	50.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.42%	
50) Toluene-d8	10.09	98	1023627	53.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.86%	
62) 4-Bromofluorobenzene	12.41	95	322675	45.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.68%	
Target Compounds						
16) Acetone	3.80	43	24299	10.220	ug/l	93
20) Methylene Chloride	4.54	84	5929m	1.208	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092019\
Data File : VN058258.D
Acq On : 20 Sep 2019 18:10
Operator : JC/SP
Sample : K4974-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB03-20190916

Manual Integrations
APPROVED

MMDadoda
9/23/2019 3:30:42 PM

Quant Time: Sep 23 05:45:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

