

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092019\
 Data File : VN058262.D
 Acq On : 20 Sep 2019 19:39
 Operator : JC/SP
 Sample : K4975-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE134D4-20190918

Quant Time: Oct 01 03:14:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	477505	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	775221	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	668124	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	274428	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	349860	52.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.22%	
35) Dibromofluoromethane	7.57	113	237645	50.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.86%	
50) Toluene-d8	10.09	98	975170	53.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.22%	
62) 4-Bromofluorobenzene	12.41	95	308781	45.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.54%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.74	101	795577	206.093	ug/l	99
16) Acetone	3.80	43	11143	4.936	ug/l	89
27) cis-1,2-Dichloroethene	6.81	96	4071	0.841	ug/l	78
44) Trichloroethene	8.83	130	90543	21.669	ug/l	97
64) Tetrachloroethene	10.63	164	35275	10.831	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN092019\
Data File : VN058262.D
Acq On : 20 Sep 2019 19:39
Operator : JC/SP
Sample : K4975-08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE134D4-20190918

Quant Time: Oct 01 03:14:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

