

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092019\
 Data File : VN058263.D
 Acq On : 20 Sep 2019 20:01
 Operator : JC/SP
 Sample : K4975-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE108D1-20190918

Quant Time: Sep 23 07:15:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	480885	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	784289	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	667019	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	269726	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	353946	52.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.70%	
35) Dibromofluoromethane	7.57	113	245081	51.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.80%	
50) Toluene-d8	10.08	98	991252	53.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.72%	
62) 4-Bromofluorobenzene	12.41	95	313305	45.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.80%	
Target Compounds						
3) Chloromethane	2.04	50	4830	0.564	ug/l	93
9) 1,1,2-Trichlorotrifluoroet	3.74	101	2592	0.667	ug/l #	92
44) Trichloroethene	8.82	130	164986	39.028	ug/l	97
64) Tetrachloroethene	10.63	164	7887	2.426	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092019\
Data File : VN058263.D
Acq On : 20 Sep 2019 20:01
Operator : JC/SP
Sample : K4975-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE108D1-20190918

Quant Time: Sep 23 07:15:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

