

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092019\
 Data File : VN058268.D
 Acq On : 20 Sep 2019 21:52
 Operator : JC/SP
 Sample : K4975-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE120D3-20190917

Quant Time: Sep 21 04:36:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	492300	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	798524	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	681678	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	265820	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	359639	52.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.90%	
35) Dibromofluoromethane	7.58	113	250979	51.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.40%	
50) Toluene-d8	10.08	98	1020245	53.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.88%	
62) 4-Bromofluorobenzene	12.41	95	324155	46.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.28%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.74	101	19276	4.843	ug/l	99
12) 1,1-Dichloroethene	3.71	96	3005	0.644	ug/l	92
27) cis-1,2-Dichloroethene	6.81	96	4097	0.821	ug/l	86
44) Trichloroethene	8.82	130	844074	196.108	ug/l	99
64) Tetrachloroethene	10.63	164	3820	1.150	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092019\
Data File : VN058268.D
Acq On : 20 Sep 2019 21:52
Operator : JC/SP
Sample : K4975-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE120D3-20190917

Quant Time: Sep 21 04:36:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

