

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092021\
 Data File : VN068560.D
 Acq On : 20 Sep 2021 14:46
 Operator : JC/MD
 Sample : VN0920MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0920MBL01

Quant Time: Sep 21 07:28:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	471137	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.970	114	681255	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	577976	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	211152	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.439	65	230264	53.280	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.560%
35) Dibromofluoromethane	8.024	113	209845	51.071	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.140%
50) Toluene-d8	10.443	98	787992	46.991	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	93.980%
62) 4-Bromofluorobenzene	12.736	95	221872	40.777	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	81.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092021\
Data File : VN068560.D
Acq On : 20 Sep 2021 14:46
Operator : JC/MD
Sample : VN0920MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0920MBL01

Quant Time: Sep 21 07:28:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

