

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092021\
 Data File : VN068565.D
 Acq On : 20 Sep 2021 19:35
 Operator : JC/MD
 Sample : M3781-02DL 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-180D2-20210915DL

Quant Time: Sep 21 07:29:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	463179	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	664609	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	574244	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	193677	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	231009	54.371	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.740%	
35) Dibromofluoromethane	8.027	113	211213	52.692	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.380%	
50) Toluene-d8	10.443	98	783247	47.811	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 95.620%	
62) 4-Bromofluorobenzene	12.736	95	212472	40.104	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 80.200%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.218	101	24020	6.043	ug/l	96
12) 1,1-Dichloroethene	4.189	96	14206	3.904	ug/l	90
32) 1,1,1-Trichloroethane	8.024	97	2980	0.400	ug/l #	43
44) Trichloroethene	9.220	130	373455	71.916	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092021\
Data File : VN068565.D
Acq On : 20 Sep 2021 19:35
Operator : JC/MD
Sample : M3781-02DL 5X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-180D2-20210915DL

Quant Time: Sep 21 07:29:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

