

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092021\
 Data File : VN068569.D
 Acq On : 20 Sep 2021 21:16
 Operator : JC/MD
 Sample : M3781-16DL 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE134D3-20210916DL

Quant Time: Sep 21 07:30:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	410433	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	602175	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	554904	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	204002	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	218106	57.931	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	115.860%
35) Dibromofluoromethane	8.024	113	201972	55.611	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	111.220%
50) Toluene-d8	10.446	98	762752	51.120	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.240%
62) 4-Bromofluorobenzene	12.733	95	216012	44.493	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.980%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.221	101	184561	52.399	ug/l	95
44) Trichloroethene	9.220	130	76208	16.197	ug/l	93
64) Tetrachloroethene	10.979	164	18661	3.794	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092021\
Data File : VN068569.D
Acq On : 20 Sep 2021 21:16
Operator : JC/MD
Sample : M3781-16DL 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE134D3-20210916DL

Quant Time: Sep 21 07:30:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

