

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092021\
 Data File : VN068571.D
 Acq On : 20 Sep 2021 22:07
 Operator : JC/MD
 Sample : M3781-05RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE135D2-20210915RE

Quant Time: Sep 21 07:31:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	398090	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	611250	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	544103	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	180689	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	229590	62.872	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 125.740%	
35) Dibromofluoromethane	8.024	113	202331	54.882	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 109.760%	
50) Toluene-d8	10.446	98	767913	50.731	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 101.460%	
62) 4-Bromofluorobenzene	12.734	95	210494	42.879	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 85.760%	
Target Compounds						
45) 1,2-Dichloropropane	9.491	63	9620	2.671	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092021\
Data File : VN068571.D
Acq On : 20 Sep 2021 22:07
Operator : JC/MD
Sample : M3781-05RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE135D2-20210915RE

Quant Time: Sep 21 07:31:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

