

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092021\
 Data File : VN068573.D
 Acq On : 20 Sep 2021 22:57
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 21 07:31:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	80	0.00
2 T	Dichlorodifluoromethane	0.408	0.324	20.6	65	0.00
3 P	Chloromethane	0.409	0.342	16.4	74	0.00
4 C	Vinyl Chloride	0.619	0.532	14.1#	73	0.00
5 T	Bromomethane	0.602	0.431	28.4#	66	0.00
6 T	Chloroethane	0.456	0.391	14.3	75	0.00
7 T	Trichlorofluoromethane	0.776	0.726	6.4	80	0.00
8 T	Diethyl Ether	0.234	0.246	-5.1	86	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.429	0.398	7.2	77	0.00
10 T	Methyl Iodide	0.612	0.569	7.0	71	0.00
11 T	Tert butyl alcohol	0.074	0.062	16.2	77	0.00
12 CM	1,1-Dichloroethene	0.393	0.373	5.1#	77	0.00
13 T	Acrolein	0.034	0.023	32.4#	57	0.00
14 T	Allyl chloride	0.499	0.517	-3.6	84	0.00
15 T	Acrylonitrile	0.175	0.191	-9.1	89	0.00
16 T	Acetone	0.152	0.138	9.2	83	0.00
17 T	Carbon Disulfide	1.055	0.848	19.6	67	0.00
18 T	Methyl Acetate	0.484	0.459	5.2	90	0.00
19 T	Methyl tert-butyl Ether	1.316	1.383	-5.1	83	0.00
20 T	Methylene Chloride	0.522	0.461	11.7	84	0.00
21 T	trans-1,2-Dichloroethene	0.448	0.418	6.7	76	0.00
22 T	Diisopropyl ether	1.080	1.225	-13.4	88	0.00
23 T	Vinyl Acetate	0.807	0.910	-12.8	87	0.00
24 P	1,1-Dichloroethane	0.732	0.738	-0.8	83	0.00
25 T	2-Butanone	0.223	0.237	-6.3	87	0.00
26 T	2,2-Dichloropropane	0.676	0.572	15.4	71	0.00
27 T	cis-1,2-Dichloroethene	0.527	0.519	1.5	80	0.00
28 T	Bromochloromethane	0.285	0.294	-3.2	77	0.00
29 T	Tetrahydrofuran	0.145	0.158	-9.0	85	0.00
30 C	Chloroform	0.866	0.863	0.3#	81	0.00
31 T	Cyclohexane	0.695	0.623	10.4	75	0.00
32 T	1,1,1-Trichloroethane	0.803	0.813	-1.2	80	0.00
33 S	1,2-Dichloroethane-d4	0.459	0.497	-8.3	86	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	81	0.00
35 S	Dibromofluoromethane	0.302	0.330	-9.3	84	0.00
36 T	1,1-Dichloropropene	0.417	0.408	2.2	76	0.00
37 T	Ethyl Acetate	0.317	0.350	-10.4	86	0.00
38 T	Carbon Tetrachloride	0.499	0.515	-3.2	78	0.00
39 T	Methylcyclohexane	0.527	0.502	4.7	72	0.00
40 TM	Benzene	1.234	1.295	-4.9	81	0.00
41 T	Methacrylonitrile	0.156	0.171	-9.6	87	0.00
42 TM	1,2-Dichloroethane	0.428	0.435	-1.6	82	0.00
43 T	Isopropyl Acetate	0.559	0.623	-11.4	88	0.00
44 TM	Trichloroethene	0.391	0.376	3.8	75	0.00
45 C	1,2-Dichloropropane	0.295	0.314	-6.4#	84	0.00
46 T	Dibromomethane	0.237	0.244	-3.0	80	0.00
47 T	Bromodichloromethane	0.450	0.494	-9.8	84	0.00
48 T	Methyl methacrylate	0.234	0.259	-10.7	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092021\
 Data File : VN068573.D
 Acq On : 20 Sep 2021 22:57
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 21 07:31:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.005	0.005	0.0	75	0.00
50 S	Toluene-d8	1.095	1.213	-10.8	82	0.00
51 T	4-Methyl-2-Pentanone	0.319	0.378	-18.5	90	0.00
52 CM	Toluene	0.819	0.893	-9.0#	81	0.00
53 T	t-1,3-Dichloropropene	0.431	0.484	-12.3	82	0.00
54 T	cis-1,3-Dichloropropene	0.472	0.513	-8.7	80	0.00
55 T	1,1,2-Trichloroethane	0.325	0.358	-10.2	82	0.00
56 T	Ethyl methacrylate	0.404	0.490	-21.3	85	0.00
57 T	1,3-Dichloropropane	0.501	0.551	-10.0	83	0.00
58 T	2-Chloroethyl Vinyl ether	0.030	0.046	-53.3#	93	0.00
59 T	2-Hexanone	0.214	0.257	-20.1	89	0.00
60 T	Dibromochloromethane	0.382	0.438	-14.7	80	0.00
61 T	1,2-Dibromoethane	0.344	0.375	-9.0	79	0.00
62 S	4-Bromofluorobenzene	0.351	0.418	-19.1	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	82	0.00
64 T	Tetrachloroethene	0.443	0.418	5.6	73	0.00
65 PM	Chlorobenzene	1.027	1.041	-1.4	80	0.00
66 T	1,1,1,2-Tetrachloroethane	0.404	0.426	-5.4	81	0.00
67 C	Ethyl Benzene	1.681	1.782	-6.0#	82	0.00
68 T	m/p-Xylenes	0.677	0.724	-6.9	80	0.00
69 T	o-Xylene	0.660	0.721	-9.2	82	0.00
70 T	Styrene	1.000	1.154	-15.4	83	0.00
71 P	Bromoform	0.310	0.359	-15.8	79	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
73 T	Isopropylbenzene	3.557	3.584	-0.8	81	0.00
74 T	N-amyl acetate	0.793	0.915	-15.4	91	0.00
75 P	1,1,2,2-Tetrachloroethane	0.997	0.996	0.1	87	0.00
76 T	1,2,3-Trichloropropane	0.789	0.816	-3.4	99	0.00
77 T	Bromobenzene	0.984	0.956	2.8	78	0.00
78 T	n-propylbenzene	3.756	3.867	-3.0	82	0.00
79 T	2-Chlorotoluene	2.315	2.344	-1.3	85	0.00
80 T	1,3,5-Trimethylbenzene	2.889	3.010	-4.2	83	0.00
81 T	trans-1,4-Dichloro-2-butene	0.226	0.260	-15.0	89	0.00
82 T	4-Chlorotoluene	2.233	2.238	-0.2	82	0.00
83 T	tert-Butylbenzene	2.638	2.751	-4.3	82	0.00
84 T	1,2,4-Trimethylbenzene	2.771	2.941	-6.1	83	0.00
85 T	sec-Butylbenzene	3.478	3.655	-5.1	82	0.00
86 T	p-Isopropyltoluene	2.936	3.073	-4.7	79	0.00
87 T	1,3-Dichlorobenzene	1.707	1.690	1.0	77	0.00
88 T	1,4-Dichlorobenzene	1.682	1.644	2.3	78	0.00
89 T	n-Butylbenzene	2.139	2.178	-1.8	79	0.00
90 T	Hexachloroethane	0.515	0.586	-13.8	89	0.00
91 T	1,2-Dichlorobenzene	1.647	1.643	0.2	79	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.158	0.159	-0.6	78	0.00
93 T	1,2,4-Trichlorobenzene	0.752	0.691	8.1	62	0.00
94 T	Hexachlorobutadiene	0.486	0.450	7.4	68	0.00
95 T	Naphthalene	1.536	1.378	10.3	58	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092021\
Data File : VN068573.D
Acq On : 20 Sep 2021 22:57
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Sep 21 07:31:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.718	0.636	11.4	60	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6