

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092024\
 Data File : VN084094.D
 Acq On : 21 Sep 2024 06:35
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Sep 21 07:27:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	132203	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	241768	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	210142	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	100218	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.577	65	108321	54.339	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.680%	
35) Dibromofluoromethane	8.165	113	78594	50.779	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.560%	
50) Toluene-d8	10.565	98	277490	51.860	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.720%	
62) 4-Bromofluorobenzene	12.847	95	106497	48.242	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 96.480%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.124	85	71085	46.299	ug/l	94
3) Chloromethane	2.359	50	77930	49.694	ug/l	97
4) Vinyl Chloride	2.512	62	80294	47.713	ug/l	100
5) Bromomethane	2.942	94	46702	48.334	ug/l	96
6) Chloroethane	3.118	64	52867	44.785	ug/l	98
7) Trichlorofluoromethane	3.500	101	137137	48.008	ug/l	100
8) Diethyl Ether	3.959	74	50541	52.782	ug/l	92
9) 1,1,2-Trichlorotrifluo...	4.371	101	75629	49.950	ug/l	99
10) Methyl Iodide	4.589	142	94794	49.227	ug/l	98
11) Tert butyl alcohol	5.518	59	89948	336.960	ug/l	99
12) 1,1-Dichloroethene	4.336	96	71384	49.922	ug/l	97
13) Acrolein	4.177	56	25976	98.625	ug/l	95
14) Allyl chloride	5.018	41	123310	48.397	ug/l	97
15) Acrylonitrile	5.718	53	220467	303.696	ug/l	98
16) Acetone	4.424	43	194428	243.303	ug/l	98
17) Carbon Disulfide	4.712	76	176094	44.025	ug/l	97
18) Methyl Acetate	5.024	43	129443	78.813	ug/l	98
19) Methyl tert-butyl Ether	5.794	73	289595	55.968	ug/l	99
20) Methylene Chloride	5.283	84	88141	53.899	ug/l	97
21) trans-1,2-Dichloroethene	5.783	96	75550	49.301	ug/l	88
22) Diisopropyl ether	6.671	45	309147	58.942	ug/l	97
23) Vinyl Acetate	6.606	43	1078596	280.814	ug/l	98
24) 1,1-Dichloroethane	6.571	63	165682	54.830	ug/l	99
25) 2-Butanone	7.483	43	305293	290.668	ug/l	99
26) 2,2-Dichloropropane	7.488	77	86054	30.924	ug/l	96
27) cis-1,2-Dichloroethene	7.483	96	97636	53.672	ug/l	98
28) Bromochloromethane	7.812	49	71135	55.901	ug/l	94
29) Tetrahydrofuran	7.841	42	195980	315.566	ug/l	97
30) Chloroform	7.965	83	173475	54.076	ug/l	99
31) Cyclohexane	8.253	56	129457	44.934	ug/l	97
32) 1,1,1-Trichloroethane	8.171	97	155837	52.830	ug/l	97
36) 1,1-Dichloropropene	8.371	75	113939	48.909	ug/l	99
37) Ethyl Acetate	7.559	43	120170	56.829	ug/l	96
38) Carbon Tetrachloride	8.365	117	131958	49.929	ug/l	99
39) Methylcyclohexane	9.600	83	120291	45.364	ug/l	98
40) Benzene	8.606	78	356299	50.814	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092024\
 Data File : VN084094.D
 Acq On : 21 Sep 2024 06:35
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Sep 21 07:27:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	72144	57.683	ug/l	99
42) 1,2-Dichloroethane	8.671	62	136544	52.114	ug/l	99
43) Isopropyl Acetate	8.688	43	233088	60.828	ug/l	98
44) Trichloroethene	9.347	130	81355	50.103	ug/l	94
45) 1,2-Dichloropropane	9.624	63	91370	53.914	ug/l	98
46) Dibromomethane	9.706	93	63281	53.310	ug/l	97
47) Bromodichloromethane	9.888	83	140802	54.808	ug/l	99
48) Methyl methacrylate	9.682	41	106372	58.510	ug/l	98
49) 1,4-Dioxane	9.694	88	34689	1159.874	ug/l #	96
51) 4-Methyl-2-Pentanone	10.447	43	647796	310.557	ug/l	97
52) Toluene	10.629	92	222059	51.775	ug/l	97
53) t-1,3-Dichloropropene	10.835	75	131667	50.564	ug/l	98
54) cis-1,3-Dichloropropene	10.312	75	135944	49.700	ug/l	98
55) 1,1,2-Trichloroethane	11.018	97	85585	53.451	ug/l	97
56) Ethyl methacrylate	10.871	69	147657	59.938	ug/l	99
57) 1,3-Dichloropropane	11.165	76	150258	54.601	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	294919	332.246	ug/l	96
59) 2-Hexanone	11.194	43	471261	305.418	ug/l	98
60) Dibromochloromethane	11.359	129	101906	56.869	ug/l	100
61) 1,2-Dibromoethane	11.465	107	84864	53.144	ug/l	99
64) Tetrachloroethene	11.100	164	72618	51.963	ug/l	95
65) Chlorobenzene	11.888	112	237072	50.759	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.959	131	86570	53.467	ug/l	99
67) Ethyl Benzene	11.965	91	435315	52.468	ug/l	100
68) m/p-Xylenes	12.070	106	321594	104.403	ug/l	99
69) o-Xylene	12.394	106	157320	52.757	ug/l	100
70) Styrene	12.412	104	272638	55.710	ug/l	100
71) Bromoform	12.576	173	65219	59.775	ug/l #	99
73) Isopropylbenzene	12.694	105	411337	53.052	ug/l	99
74) N-amyl acetate	12.494	43	193056	58.359	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.935	83	125103	54.915	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	160189	79.215	ug/l	63
77) Bromobenzene	12.976	156	95664	51.257	ug/l	100
78) n-propylbenzene	13.035	91	485551	53.823	ug/l	99
79) 2-Chlorotoluene	13.123	91	300892	52.299	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	349159	53.413	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	37215	48.199	ug/l	98
82) 4-Chlorotoluene	13.217	91	302580	52.408	ug/l	100
83) tert-Butylbenzene	13.435	119	301256	52.675	ug/l	100
84) 1,2,4-Trimethylbenzene	13.482	105	351934	53.596	ug/l	100
85) sec-Butylbenzene	13.617	105	413619	54.533	ug/l	99
86) p-Isopropyltoluene	13.729	119	338580	53.252	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	176720	50.172	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	177838	49.863	ug/l	98
89) n-Butylbenzene	14.053	91	295380	51.743	ug/l	99
90) Hexachloroethane	14.335	117	65856	51.527	ug/l	98
91) 1,2-Dichlorobenzene	14.106	146	175798	51.716	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.717	75	26448	58.239	ug/l	94
93) 1,2,4-Trichlorobenzene	15.388	180	84406	46.271	ug/l	99
94) Hexachlorobutadiene	15.500	225	37509	48.419	ug/l	97
95) Naphthalene	15.635	128	290746	52.389	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	83388	47.832	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092024\
Data File : VN084094.D
Acq On : 21 Sep 2024 06:35
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Sep 21 07:27:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092024\
 Data File : VN084094.D
 Acq On : 21 Sep 2024 06:35
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :


Quant Time: Sep 21 07:27:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:168 Resp: 132203
Ion Ratio Lower Upper
168 100
99 69.6 54.2 81.2



#2
Dichlorodifluoromethane
Concen: 46.299 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion: 85 Resp: 71085
Ion Ratio Lower Upper
85 100
87 34.5 15.7 47.0









#3
Chloromethane
Concen: 49.694 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 50 Resp: 77930
Ion Ratio Lower Upper
50 100
52 32.4 27.2 40.8



#4
Vinyl Chloride
Concen: 47.713 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion: 62 Resp: 80294
Ion Ratio Lower Upper
62 100
64 30.8 24.7 37.1









#5
Bromomethane
Concen: 48.334 ug/l
RT: 2.942 min Scan# 1
Delta R.T. 0.035 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 94 Resp: 46702
Ion Ratio Lower Upper
94 100
96 91.2 76.0 114.0



#6
Chloroethane
Concen: 44.785 ug/l
RT: 3.118 min Scan# 198
Delta R.T. 0.035 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion: 64 Resp: 52867
Ion Ratio Lower Upper
64 100
66 30.9 23.9 35.9









#7
 Trichlorofluoromethane
 Concen: 48.008 ug/l
 RT: 3.500 min Scan# 20
 Delta R.T. 0.029 min
 Lab File: VN084094.D
 Acq: 21 Sep 2024 06:35




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:101 Resp: 137137
 Ion Ratio Lower Upper
 101 100
 103 64.0 51.4 77.0



#8
 Diethyl Ether
 Concen: 52.782 ug/l
 RT: 3.959 min Scan# 341
 Delta R.T. 0.006 min
 Lab File: VN084094.D
 Acq: 21 Sep 2024 06:35

Tgt Ion: 74 Resp: 50541
 Ion Ratio Lower Upper
 74 100
 45 101.2 46.7 140.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 49.950 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.018 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:101 Resp: 75629
Ion Ratio Lower Upper
101 100
85 46.7 36.2 54.4
151 71.8 57.7 86.5



#10
Methyl Iodide
Concen: 49.227 ug/l
RT: 4.589 min Scan# 448
Delta R.T. 0.012 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion:142 Resp: 94794
Ion Ratio Lower Upper
142 100
127 48.1 37.5 56.3
141 15.5 11.0 16.4









#11
Tert butyl alcohol
Concen: 336.960 ug/l
RT: 5.518 min Scan# 608
Delta R.T. -0.012 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 59 Resp: 89948
Ion Ratio Lower Upper
59 100
57 10.4 8.7 13.1



#12
1,1-Dichloroethene
Concen: 49.922 ug/l
RT: 4.336 min Scan# 405
Delta R.T. 0.006 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion: 96 Resp: 71384
Ion Ratio Lower Upper
96 100
61 178.1 143.4 215.0
98 60.2 52.7 79.1





#13

Acrolein

Concen: 98.625 ug/l

RT: 4.177 min Scan# 31

Delta R.T. 0.006 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 56 Resp: 25976

Ion Ratio Lower Upper

56 100

55 76.8 58.2 87.2



#14

Allyl chloride

Concen: 48.397 ug/l

RT: 5.018 min Scan# 521

Delta R.T. -0.000 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35




Tgt Ion: 41 Resp: 123310


Ion Ratio Lower Upper

41 100

39 77.1 59.3 88.9


76 35.2 27.7 41.5








#15
Acrylonitrile
Concen: 303.696 ug/l
RT: 5.718 min Scan# 640
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 53 Resp: 220467
Ion Ratio Lower Upper
53 100
52 84.6 66.1 99.1
51 37.2 29.6 44.4



#16
Acetone
Concen: 243.303 ug/l
RT: 4.424 min Scan# 420
Delta R.T. -0.006 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion: 43 Resp: 194428
Ion Ratio Lower Upper
43 100
58 28.6 23.7 35.5





#17

Carbon Disulfide

Concen: 44.025 ug/l

RT: 4.712 min Scan# 469

Delta R.T. 0.012 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 76 Resp: 176094

Ion Ratio Lower Upper

76 100

78 8.0 7.1 10.7



#18

Methyl Acetate

Concen: 78.813 ug/l

RT: 5.024 min Scan# 522

Delta R.T. -0.000 min

Lab File: VN084094.D




Acq: 21 Sep 2024 06:35


Tgt Ion: 43 Resp: 129443

Ion Ratio Lower Upper

43 100

74 22.9 19.0 28.4





#19

Methyl tert-butyl Ether

Concen: 55.968 ug/l

RT: 5.794 min Scan# 61

Delta R.T. 0.006 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 73 Resp: 289595

Ion Ratio Lower Upper

73 100

57 21.8 17.0 25.6



#20

Methylene Chloride

Concen: 53.899 ug/l

RT: 5.283 min Scan# 566

Delta R.T. 0.012 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35

Tgt Ion: 84 Resp: 88141




Ion Ratio Lower Upper


84 100

49 121.8 93.4 140.0

51 36.8 29.2 43.8

86 57.3 46.9 70.3









#21
trans-1,2-Dichloroethene
Concen: 49.301 ug/l
RT: 5.783 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 96 Resp: 75550
Ion Ratio Lower Upper
96 100
61 150.3 106.0 159.0
98 60.0 52.2 78.2



#22
Diisopropyl ether
Concen: 58.942 ug/l
RT: 6.671 min Scan# 802
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion: 45 Resp: 309147
Ion Ratio Lower Upper
45 100
43 51.5 43.4 65.0
87 26.7 22.6 33.8
59 11.3 9.1 13.7









#23
 Vinyl Acetate
 Concen: 280.814 ug/l
 RT: 6.606 min Scan# 791
 Delta R.T. 0.006 min
 Lab File: VN084094.D
 Acq: 21 Sep 2024 06:35




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 43 Resp: 1078596
 Ion Ratio Lower Upper
 43 100
 86 10.1 8.7 13.1



#24
 1,1-Dichloroethane
 Concen: 54.830 ug/l
 RT: 6.571 min Scan# 785
 Delta R.T. 0.012 min
 Lab File: VN084094.D
 Acq: 21 Sep 2024 06:35

Tgt Ion: 63 Resp: 165682
 Ion Ratio Lower Upper
 63 100
 98 7.1 3.5 10.4
 100 4.5 2.5 7.5









#25
2-Butanone
Concen: 290.668 ug/l
RT: 7.483 min Scan# 940
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 305293
Ion Ratio Lower Upper
43 100
72 26.5 20.8 31.2



#26
2,2-Dichloropropane
Concen: 30.924 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion: 77 Resp: 86054
Ion Ratio Lower Upper
77 100
97 22.0 10.2 30.4









#27
 cis-1,2-Dichloroethene
 Concen: 53.672 ug/l
 RT: 7.483 min Scan# 940
 Delta R.T. -0.000 min
 Lab File: VN084094.D
 Acq: 21 Sep 2024 06:35




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 96 Resp: 97636
 Ion Ratio Lower Upper
 96 100
 61 149.6 0.0 306.2
 98 63.7 0.0 128.6



#28
 Bromochloromethane
 Concen: 55.901 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. -0.000 min
 Lab File: VN084094.D
 Acq: 21 Sep 2024 06:35

Tgt Ion: 49 Resp: 71135
 Ion Ratio Lower Upper
 49 100
 129 1.6 0.0 3.4
 130 70.2 60.2 90.2





#29

Tetrahydrofuran

Concen: 315.566 ug/l

RT: 7.841 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 42 Resp: 195980

Ion Ratio Lower Upper

42 100

72 42.9 35.9 53.9

71 40.2 33.9 50.9



#30

Chloroform

Concen: 54.076 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. 0.006 min

Lab File: VN084094.D




Acq: 21 Sep 2024 06:35


Tgt Ion: 83 Resp: 173475

Ion Ratio Lower Upper

83 100


85 65.4 52.8 79.2








#31
Cyclohexane
Concen: 44.934 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 56 Resp: 129457
Ion Ratio Lower Upper
56 100
69 31.4 26.9 40.3
84 83.4 69.2 103.8



#32
1,1,1-Trichloroethane
Concen: 52.830 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. 0.006 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion: 97 Resp: 155837
Ion Ratio Lower Upper
97 100
99 62.8 52.3 78.5
61 49.0 38.2 57.2





#33

1,2-Dichloroethane-d4

Concen: 54.339 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. -0.000 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 65 Resp: 108321

Ion Ratio Lower Upper

65 100

67 50.7 0.0 100.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. -0.000 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35




Tgt Ion: 114 Resp: 241768


Ion Ratio Lower Upper

114 100

63 23.3 0.0 43.4

88 16.5 0.0 32.6





#35

Dibromofluoromethane

Concen: 50.779 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN084094.D





Acq: 21 Sep 2024 06:35

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion:	113	Resp:	78594
Ion Ratio	Lower	Upper	
113	100		
111	103.5	82.6	124.0
192	16.2	14.3	21.5



#36

1,1-Dichloropropene

Concen: 48.909 ug/l




RT: 8.371 min Scan# 1091


Delta R.T. -0.000 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35

Tgt Ion:	75	Resp:	113939
Ion Ratio	Lower	Upper	
75	100		
110	32.8	16.7	50.1
77	30.6	25.2	37.8









#37
Ethyl Acetate
Concen: 56.829 ug/l
RT: 7.559 min Scan# 91
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 120170
Ion Ratio Lower Upper
43 100
61 12.2 11.1 16.7
70 9.8 8.8 13.2



#38
Carbon Tetrachloride
Concen: 49.929 ug/l
RT: 8.365 min Scan# 1090
Delta R.T. 0.006 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion: 117 Resp: 131958
Ion Ratio Lower Upper
117 100
119 94.4 76.1 114.1
121 29.9 24.1 36.1





#39

Methylcyclohexane

Concen: 45.364 ug/l

RT: 9.600 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 83 Resp: 120291

Ion Ratio Lower Upper

83 100

55 80.3 62.2 93.4

98 47.3 37.6 56.4



#40

Benzene

Concen: 50.814 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. 0.006 min

Lab File: VN084094.D




Acq: 21 Sep 2024 06:35


Tgt Ion: 78 Resp: 356299

Ion Ratio Lower Upper

78 100

77 23.4 19.0 28.6









#41
Methacrylonitrile
Concen: 57.683 ug/l
RT: 7.777 min Scan# 990
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:	41	Resp:	72144
Ion	Ratio	Lower	Upper
41	100		
39	56.6	45.4	68.2
67	68.5	55.9	83.9
52	28.9	24.1	36.1



#42
1,2-Dichloroethane
Concen: 52.114 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion:	62	Resp:	136544
Ion	Ratio	Lower	Upper
62	100		
98	7.9	0.0	16.6





#43

Isopropyl Acetate

Concen: 60.828 ug/l

RT: 8.688 min Scan# 1145

Delta R.T. -0.000 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 43 Resp: 233088

Ion Ratio Lower Upper

43 100

61 23.4 19.4 29.0

87 11.8 9.8 14.6



#44

Trichloroethene

Concen: 50.103 ug/l

RT: 9.347 min Scan# 1257

Delta R.T. -0.006 min

Lab File: VN084094.D




Acq: 21 Sep 2024 06:35


Tgt Ion:130 Resp: 81355

Ion Ratio Lower Upper

130 100

95 106.2 0.0 200.4









#45
1,2-Dichloropropane
Concen: 53.914 ug/l
RT: 9.624 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 63 Resp: 91370
Ion Ratio Lower Upper
63 100
65 31.3 25.9 38.9



#46
Dibromomethane
Concen: 53.310 ug/l
RT: 9.706 min Scan# 1318
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35


Tgt Ion: 93 Resp: 63281
Ion Ratio Lower Upper
93 100
95 83.3 65.3 97.9
174 84.2 71.0 106.4








#47
Bromodichloromethane
Concen: 54.808 ug/l
RT: 9.888 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 83 Resp: 140802
Ion Ratio Lower Upper
83 100
85 64.1 51.8 77.8
127 8.3 6.6 9.8



#48
Methyl methacrylate
Concen: 58.510 ug/l
RT: 9.682 min Scan# 1314
Delta R.T. 0.006 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion: 41 Resp: 106372
Ion Ratio Lower Upper
41 100
69 79.3 64.3 96.5
39 58.0 48.9 73.3





#49

1,4-Dioxane

Concen: 1159.874 ug/l

RT: 9.694 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 88 Resp: 34689

Ion Ratio Lower Upper

88 100

43 33.0 0.0 0.0#

58 74.3 57.1 85.7



#50

Toluene-d8

Concen: 51.860 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN084094.D




Acq: 21 Sep 2024 06:35


Tgt Ion: 98 Resp: 277490

Ion Ratio Lower Upper

98 100

100 65.9 52.0 78.0









#51
4-Methyl-2-Pentanone
Concen: 310.557 ug/l
RT: 10.447 min Scan# 1443
Delta R.T. 0.006 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 647796
Ion Ratio Lower Upper
43 100
58 37.7 31.6 47.4



#52
Toluene
Concen: 51.775 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion: 92 Resp: 222059
Ion Ratio Lower Upper
92 100
91 172.8 134.7 202.1









#53
 t-1,3-Dichloropropene
 Concen: 50.564 ug/l
 RT: 10.835 min Scan# 1510
 Delta R.T. -0.000 min
 Lab File: VN084094.D
 Acq: 21 Sep 2024 06:35




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 75 Resp: 131667
 Ion Ratio Lower Upper
 75 100
 77 33.5 25.8 38.6



#54
 cis-1,3-Dichloropropene
 Concen: 49.700 ug/l
 RT: 10.312 min Scan# 1421
 Delta R.T. -0.000 min
 Lab File: VN084094.D
 Acq: 21 Sep 2024 06:35

Tgt Ion: 75 Resp: 135944
 Ion Ratio Lower Upper
 75 100
 77 30.6 25.0 37.6
 39 54.8 42.6 64.0





#55

1,1,2-Trichloroethane

Concen: 53.451 ug/l

RT: 11.018 min Scan# 1541

Delta R.T. -0.000 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion: 97 Resp: 85585





Ion Ratio Lower Upper

97 100

83 82.4 67.5 101.3

85 56.9 44.2 66.2

99 58.7 50.1 75.1



#56

Ethyl methacrylate

Concen: 59.938 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35




Tgt Ion: 69 Resp: 147657


Ion Ratio Lower Upper

69 100

41 68.7 53.9 80.9

39 40.3 32.6 48.8





#57

1,3-Dichloropropane

Concen: 54.601 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. 0.006 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 76 Resp: 150258

Ion Ratio Lower Upper

76 100

78 32.8 25.8 38.6



#58

2-Chloroethyl Vinyl ether

Concen: 332.246 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.000 min

Lab File: VN084094.D




Acq: 21 Sep 2024 06:35


Tgt Ion: 63 Resp: 294919

Ion Ratio Lower Upper

63 100

106 27.0 23.3 34.9









#59
2-Hexanone
Concen: 305.418 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 471261
Ion Ratio Lower Upper
43 100
58 52.7 27.2 81.5



#60
Dibromochloromethane
Concen: 56.869 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion:129 Resp: 101906
Ion Ratio Lower Upper
129 100
127 77.8 39.0 116.9









#61
1,2-Dibromoethane
Concen: 53.144 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:107 Resp: 84864
Ion Ratio Lower Upper
107 100
109 93.5 75.6 113.4



#62
4-Bromofluorobenzene
Concen: 48.242 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion: 95 Resp: 106497
Ion Ratio Lower Upper
95 100
174 69.7 0.0 139.8
176 68.2 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:117 Resp: 210142
Ion Ratio Lower Upper
117 100
82 60.8 47.4 71.0
119 31.9 26.3 39.5



#64
Tetrachloroethene
Concen: 51.963 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion:164 Resp: 72618
Ion Ratio Lower Upper
164 100
166 118.1 103.7 155.5
129 98.6 79.9 119.9
131 94.1 78.4 117.6









#65
Chlorobenzene
Concen: 50.759 ug/l
RT: 11.888 min Scan# 112
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:112 Resp: 237072
Ion Ratio Lower Upper
112 100
114 32.8 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 53.467 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion:131 Resp: 86570
Ion Ratio Lower Upper
131 100
133 94.2 47.5 142.6
119 64.2 31.6 94.8









#67
Ethyl Benzene
Concen: 52.468 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 435315
Ion Ratio Lower Upper
91 100
106 30.2 24.3 36.5



#68
m/p-Xylenes
Concen: 104.403 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion:106 Resp: 321594
Ion Ratio Lower Upper
106 100
91 213.1 169.2 253.8





#69

o-Xylene

Concen: 52.757 ug/l

RT: 12.394 min Scan# 1775

Delta R.T. -0.000 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:106 Resp: 157320

Ion Ratio Lower Upper

106 100

91 225.7 112.9 338.6



#70

Styrene

Concen: 55.710 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. -0.000 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35




Tgt Ion:104 Resp: 272638


Ion Ratio Lower Upper

104 100

78 54.2 43.3 64.9

103 54.8 43.4 65.2









#71
Bromoform
Concen: 59.775 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:173 Resp: 65219
Ion Ratio Lower Upper
173 100
175 48.7 24.1 72.2
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion:152 Resp: 100218
Ion Ratio Lower Upper
152 100
115 65.2 32.2 96.6
150 175.5 0.0 348.6





#73

Isopropylbenzene

Concen: 53.052 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. -0.000 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:105 Resp: 411337

Ion Ratio Lower Upper

105 100

120 25.6 12.6 37.8



#74

N-ethyl acetate

Concen: 58.359 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. -0.000 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35

Tgt Ion: 43 Resp: 193056




Ion Ratio Lower Upper


43 100

70 41.8 34.5 51.7

55 24.7 18.6 28.0

61 23.6 19.2 28.8









#75
1,1,2,2-Tetrachloroethane
Concen: 54.915 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 125103
Ion Ratio Lower Upper
83 100
131 10.3 5.2 15.6
85 64.3 31.9 95.9



#76
1,2,3-Trichloropropane
Concen: 79.215 ug/l
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion: 75 Resp: 160189
Ion Ratio Lower Upper
75 100
77 136.6 95.2 285.6









#77
Bromobenzene
Concen: 51.257 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:156 Resp: 95664
Ion Ratio Lower Upper
156 100
77 228.7 114.3 343.1
158 96.7 48.9 146.7



#78
n-propylbenzene
Concen: 53.823 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion: 91 Resp: 485551
Ion Ratio Lower Upper
91 100
120 21.5 11.0 33.0









#79
2-Chlorotoluene
Concen: 52.299 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 300892
Ion Ratio Lower Upper
91 100
126 31.3 15.4 46.2



#80
1,3,5-Trimethylbenzene
Concen: 53.413 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion: 105 Resp: 349159
Ion Ratio Lower Upper
105 100
120 46.1 23.4 70.3









#81
trans-1,4-Dichloro-2-butene
Concen: 48.199 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 75 Resp: 37215
Ion Ratio Lower Upper
75 100
53 103.0 81.2 121.8
89 45.8 35.7 53.5



#82
4-Chlorotoluene
Concen: 52.408 ug/l
RT: 13.217 min Scan# 1915
Delta R.T. -0.006 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion: 91 Resp: 302580
Ion Ratio Lower Upper
91 100
126 31.1 15.4 46.2





#83

tert-Butylbenzene

Concen: 52.675 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:119 Resp: 301256

Ion Ratio Lower Upper

119 100

91 69.2 34.5 103.6

134 25.7 12.8 38.4



#84

1,2,4-Trimethylbenzene

Concen: 53.596 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. -0.000 min

Lab File: VN084094.D




Acq: 21 Sep 2024 06:35


Tgt Ion:105 Resp: 351934

Ion Ratio Lower Upper

105 100

120 43.5 21.8 65.4









#85
 sec-Butylbenzene
 Concen: 54.533 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. -0.000 min
 Lab File: VN084094.D
 Acq: 21 Sep 2024 06:35




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:105 Resp: 413619
 Ion Ratio Lower Upper
 105 100
 134 18.9 9.7 29.1



#86
 p-Isopropyltoluene
 Concen: 53.252 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN084094.D
 Acq: 21 Sep 2024 06:35

Tgt Ion:119 Resp: 338580
 Ion Ratio Lower Upper
 119 100
 134 25.4 12.8 38.3
 91 26.1 12.9 38.7





#87

1,3-Dichlorobenzene

Concen: 50.172 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:146 Resp: 176720

Ion Ratio Lower Upper

146 100

111 44.1 21.7 65.1

148 63.2 31.7 95.1



#88

1,4-Dichlorobenzene

Concen: 49.863 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35




Tgt Ion:146 Resp: 177838


Ion Ratio Lower Upper

146 100

111 41.6 21.3 64.0


148 62.4 32.0 96.2








#89
n-Butylbenzene
Concen: 51.743 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 91 Resp: 295380
Ion Ratio Lower Upper
91 100
92 52.6 26.7 80.1
134 23.7 12.2 36.4



#90
Hexachloroethane
Concen: 51.527 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. 0.006 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35


Tgt Ion: 117 Resp: 65856
Ion Ratio Lower Upper
117 100
201 66.9 32.8 98.3








#91
1,2-Dichlorobenzene
Concen: 51.716 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:146 Resp: 175798
Ion Ratio Lower Upper
146 100
111 43.6 22.8 68.3
148 63.0 32.1 96.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 58.239 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN084094.D
Acq: 21 Sep 2024 06:35

Tgt Ion: 75 Resp: 26448
Ion Ratio Lower Upper
75 100
155 68.6 35.9 107.7
157 88.1 47.5 142.5





#93

1,2,4-Trichlorobenzene

Concen: 46.271 ug/l

RT: 15.388 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:180 Resp: 84406

Ion Ratio Lower Upper

180 100

182 94.9 48.1 144.3

145 33.6 16.9 50.7



#94

Hexachlorobutadiene

Concen: 48.419 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35




Tgt Ion:225 Resp: 37509


Ion Ratio Lower Upper

225 100

223 61.5 31.9 95.5

227 62.0 32.0 96.2





#95

Naphthalene

Concen: 52.389 ug/l

RT: 15.635 min Scan# 2326

Delta R.T. -0.000 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:128 Resp: 290746

Ion Ratio Lower Upper

128 100

127 12.8 10.2 15.2

129 10.6 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 47.832 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. -0.006 min

Lab File: VN084094.D

Acq: 21 Sep 2024 06:35

Tgt Ion:180 Resp: 83388

Ion Ratio Lower Upper

180 100

182 95.4 47.1 141.4

145 37.7 17.8 53.5

