

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN092024\  
 Data File : VN084061.D  
 Acq On : 20 Sep 2024 11:08  
 Operator : JC\MD  
 Sample : VN0920WBL01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0920WBL01

Quant Time: Sep 20 23:23:56 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N090624W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Fri Sep 06 23:42:47 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Bromochloromethane       | 7.812  | 128   | 26514    | 30.000   | ug/l  | # 0.00   |
| 28) 1,4-Difluorobenzene     | 9.100  | 114   | 145231   | 30.000   | ug/l  | 0.00     |
| 57) Chlorobenzene-d5        | 11.865 | 117   | 128164   | 30.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 27) 1,2-Dichloroethane-d4   | 8.577  | 65    | 75601    | 31.194   | ug/l  | 0.00     |
| Spiked Amount               | 30.000 | Range | 91 - 110 | Recovery | =     | 103.967% |
| 60) 4-Bromofluorobenzene    | 12.853 | 95    | 56304    | 25.490   | ug/l  | 0.00     |
| Spiked Amount               | 30.000 | Range | 63 - 112 | Recovery | =     | 84.967%  |
| 63) Toluene-d8              | 10.565 | 98    | 180916   | 30.486   | ug/l  | 0.00     |
| Spiked Amount               | 30.000 | Range | 91 - 112 | Recovery | =     | 101.633% |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN092024\  
Data File : VN084061.D  
Acq On : 20 Sep 2024 11:08  
Operator : JC\MD  
Sample : VN0920WBL01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 6 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0920WBL01

Quant Time: Sep 20 23:23:56 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N090624W.M  
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
QLast Update : Fri Sep 06 23:42:47 2024  
Response via : Initial Calibration

