

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092024\
 Data File : VN084067.D
 Acq On : 20 Sep 2024 13:56
 Operator : JC\MD
 Sample : P4093-02 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OUTFALL-DSN-002

Quant Time: Sep 20 23:26:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Sep 06 23:42:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.818	128	24206	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	123476	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	110994	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	66364	29.994	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.967%
60) 4-Bromofluorobenzene	12.847	95	47799	24.987	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	83.300%
63) Toluene-d8	10.565	98	155881	30.330	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.100%
Target Compounds						
15) Acetone	4.436	58	4521	18.562	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092024\
Data File : VN084067.D
Acq On : 20 Sep 2024 13:56
Operator : JC\MD
Sample : P4093-02 20X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OUTFALL-DSN-002

Quant Time: Sep 20 23:26:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090624W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Sep 06 23:42:47 2024
Response via : Initial Calibration

