

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092024\
 Data File : VN084072.D
 Acq On : 20 Sep 2024 15:57
 Operator : JC\MD
 Sample : PB163512ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB163512ZHE#05

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/23/2024
 Supervised By : Mahesh Dadoda 09/23/2024

Quant Time: Sep 23 01:04:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Sep 06 23:42:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.812	128	23249	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	127735	30.000	ug/l	# 0.00
57) Chlorobenzene-d5	11.865	117	119096	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.577	65	64615	30.405	ug/l	0.00
Spiked Amount 30.000	Range	91 - 110	Recovery	=	101.367%	
60) 4-Bromofluorobenzene	12.847	95	48630	23.692	ug/l	0.00
Spiked Amount 30.000	Range	63 - 112	Recovery	=	78.967%	
63) Toluene-d8	10.571	98	162988	29.556	ug/l	0.00
Spiked Amount 30.000	Range	91 - 112	Recovery	=	98.533%	
Target Compounds						
15) Acetone	4.436	58	13652	58.358	ug/l	Qvalue 98
18) Methylene Chloride	5.271	84	5067	3.205	ug/l	# 85
44) Isopropyl Acetate	8.688	43	194935m	58.630	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092024\
Data File : VN084072.D
Acq On : 20 Sep 2024 15:57
Operator : JC\MD
Sample : PB163512ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB163512ZHE#05

Manual Integrations
APPROVED

Reviewed By : John Carlone 09/23/2024
Supervised By : Mahesh Dadoda 09/23/2024

Quant Time: Sep 23 01:04:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090624W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Sep 06 23:42:47 2024
Response via : Initial Calibration

