

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092024\
 Data File : VN084081.D
 Acq On : 21 Sep 2024 01:23
 Operator : JC\MD
 Sample : VN0920WBL03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0920WBL03

Quant Time: Sep 23 07:22:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	152902	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	262064	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	223618	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	93481	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	120766	52.381	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.760%	
35) Dibromofluoromethane	8.165	113	86299	51.439	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.880%	
50) Toluene-d8	10.565	98	318194	54.862	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 109.720%	
62) 4-Bromofluorobenzene	12.847	95	108507	45.345	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 90.700%	
Target Compounds						
3) Chloromethane	2.359	50	267	Below Cal	#	41
18) Methyl Acetate	5.024	43	184	Below Cal	#	52

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092024\
Data File : VN084081.D
Acq On : 21 Sep 2024 01:23
Operator : JC\MD
Sample : VN0920WBL03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0920WBL03

Quant Time: Sep 23 07:22:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

