

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092024\
 Data File : VN084085.D
 Acq On : 21 Sep 2024 02:59
 Operator : JC\MD
 Sample : P4103-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-22A

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/23/2024
 Supervised By : Mahesh Dadoda 09/23/2024

Quant Time: Sep 21 03:54:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	157983	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	276231	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	241651	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	94833	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	122054	51.237	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 102.480%			
35) Dibromofluoromethane	8.165	113	88946	50.298	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.600%			
50) Toluene-d8	10.565	98	338075	55.300	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 110.600%			
62) 4-Bromofluorobenzene	12.847	95	116432	46.162	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 92.320%			
Target Compounds					Qvalue	
16) Acetone	4.430	43	49221	51.543	ug/l	97
20) Methylene Chloride	5.265	84	3932	2.012	ug/l #	88
43) Isopropyl Acetate	8.688	43	110441m	25.225	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092024\
Data File : VN084085.D
Acq On : 21 Sep 2024 02:59
Operator : JC\MD
Sample : P4103-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-22A

Quant Time: Sep 21 03:54:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 09/23/2024
Supervised By : Mahesh Dadoda 09/23/2024

