

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092024\
 Data File : VN084089.D
 Acq On : 21 Sep 2024 04:35
 Operator : JC\MD
 Sample : P4103-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14-7.5-10B

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/23/2024
 Supervised By : Mahesh Dadoda 09/23/2024

Quant Time: Sep 21 05:56:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	165247	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	284466	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	248483	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	108316	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	125229	50.259	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 100.520%			
35) Dibromofluoromethane	8.165	113	90111	49.481	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.960%			
50) Toluene-d8	10.565	98	347258	55.158	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 110.320%			
62) 4-Bromofluorobenzene	12.847	95	125607	48.358	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 96.720%			
Target Compounds						
16) Acetone	4.430	43	53582	53.643	ug/l	96
20) Methylene Chloride	5.277	84	4430	2.167	ug/l #	76
25) 2-Butanone	7.494	43	13470	10.260	ug/l	95
43) Isopropyl Acetate	8.688	43	130215m	28.881	ug/l	
52) Toluene	10.623	92	10723	2.125	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092024\
Data File : VN084089.D
Acq On : 21 Sep 2024 04:35
Operator : JC\MD
Sample : P4103-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14-7.5-10B

Quant Time: Sep 21 05:56:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 09/23/2024
Supervised By :Mahesh Dadoda 09/23/2024

