

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092118\
 Data File : VN051360.D
 Acq On : 21 Sep 2018 10:54
 Operator : MD\SY
 Sample : VN0921WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0921WBL01

Quant Time: Sep 21 13:53:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N082818W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 29 01:38:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	118877	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	618561	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	518568	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	213558	28.15	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	93.83%
60) 4-Bromofluorobenzene	12.40	95	168814	21.89	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	72.97%
63) Toluene-d8	10.09	98	698300	28.88	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	96.27%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092118\
Data File : VN051360.D
Acq On : 21 Sep 2018 10:54
Operator : MD\SY
Sample : VN0921WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0921WBL01

Quant Time: Sep 21 13:53:45 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N082818W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 29 01:38:44 2018
Response via : Initial Calibration

