

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092118\
 Data File : VN051362.D
 Acq On : 21 Sep 2018 11:54
 Operator : MD\SY
 Sample : J5015-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPDES-1

Quant Time: Sep 21 14:04:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N082818W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 29 01:38:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	114854	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	582457	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	486949	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	206956	28.24	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	94.13%
60) 4-Bromofluorobenzene	12.40	95	158554	21.90	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	73.00%
63) Toluene-d8	10.09	98	656812	28.92	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	96.40%

Target Compounds					Qvalue
25) Chloroform	7.37	83	11376	0.917 ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN092118\
Data File : VN051362.D
Acq On : 21 Sep 2018 11:54
Operator : MD\SY
Sample : J5015-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPDES-1

Quant Time: Sep 21 14:04:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\624N082818W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 29 01:38:44 2018
Response via : Initial Calibration

