

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092120\
 Data File : VN063493.D
 Acq On : 21 Sep 2020 18:08
 Operator : JC/MD
 Sample : L4057-01ME
 Misc : 1.10g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JA-DRUM-COMPME

Quant Time: Sep 22 04:11:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	245353	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	458042	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	450208	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	200782	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	195179	46.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.46%	
35) Dibromofluoromethane	7.55	113	138334	43.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.26%	
50) Toluene-d8	10.06	98	580388	48.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.36%	
62) 4-Bromofluorobenzene	12.38	95	228166	52.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.40%	
Target Compounds						
39) Methylcyclohexane	9.05	83	8118	1.533	ug/l	93
52) Toluene	10.13	92	8866	1.059	ug/l	98
68) m/p-Xylenes	11.59	106	13809	2.128	ug/l	99
84) 1,2,4-Trimethylbenzene	13.02	105	17641	1.302	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092120\
Data File : VN063493.D
Acq On : 21 Sep 2020 18:08
Operator : JC/MD
Sample : L4057-01ME
Misc : 1.10g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JA-DRUM-COMPME

Quant Time: Sep 22 04:11:13 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

