

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092120\
 Data File : VN063496.D
 Acq On : 21 Sep 2020 19:19
 Operator : JC/MD
 Sample : PB131742TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131742TB

Manual Integrations
 APPROVED

MMDadoda
 9/22/2020 2:07:19 PM

Quant Time: Sep 22 04:25:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	272936	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	517800	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	522786	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	228739	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	218922	47.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.24%	
35) Dibromofluoromethane	7.55	113	162255	45.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.52%	
50) Toluene-d8	10.06	98	662408	49.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.28%	
62) 4-Bromofluorobenzene	12.38	95	271754	55.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.00%	
Target Compounds						
16) Acetone	3.77	43	59940	40.143	ug/l	96
18) Methyl Acetate	4.29	43	4878m	1.287	ug/l	
20) Methylene Chloride	4.51	84	34221	9.046	ug/l	94
43) Isopropyl Acetate	8.13	43	92972	9.604	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092120\
Data File : VN063496.D
Acq On : 21 Sep 2020 19:19
Operator : JC/MD
Sample : PB131742TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131742TB

Manual Integrations
APPROVED

MMDadoda
9/22/2020 2:07:19 PM

Quant Time: Sep 22 04:25:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

