

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092120\
 Data File : VN063498.D
 Acq On : 21 Sep 2020 20:07
 Operator : JC/MD
 Sample : L4072-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-9

Quant Time: Sep 22 04:30:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	261844	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	498190	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	487235	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	210760	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	213350	47.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.74%	
35) Dibromofluoromethane	7.55	113	157471	45.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.32%	
50) Toluene-d8	10.06	98	634390	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
62) 4-Bromofluorobenzene	12.38	95	251036	52.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.62%	
Target Compounds						
16) Acetone	3.78	43	6465	4.513	ug/l #	Qvalue 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092120\
Data File : VN063498.D
Acq On : 21 Sep 2020 20:07
Operator : JC/MD
Sample : L4072-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-9

Quant Time: Sep 22 04:30:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

