

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092120\
 Data File : VN063503.D
 Acq On : 21 Sep 2020 23:17
 Operator : JC/MD
 Sample : VN0921WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0921WBL02

Quant Time: Sep 22 05:14:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	244866	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	457680	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	449584	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	195293	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	200120	48.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.02%	
35) Dibromofluoromethane	7.55	113	147273	46.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.96%	
50) Toluene-d8	10.06	98	579390	48.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.26%	
62) 4-Bromofluorobenzene	12.38	95	232923	53.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.66%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092120\
Data File : VN063503.D
Acq On : 21 Sep 2020 23:17
Operator : JC/MD
Sample : VN0921WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0921WBL02

Quant Time: Sep 22 05:14:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

