

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092120\
 Data File : VN063515.D
 Acq On : 22 Sep 2020 4:06
 Operator : JC/MD
 Sample : L4082-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE134D3-20200915

Quant Time: Sep 22 09:33:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	229872	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	428663	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	432901	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	180916	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	186302	47.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.22%	
35) Dibromofluoromethane	7.55	113	139233	46.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.84%	
50) Toluene-d8	10.06	98	550224	49.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.62%	
62) 4-Bromofluorobenzene	12.38	95	217499	53.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.34%	
Target Compounds						
					Qvalue	
3) Chloromethane	2.04	50	1914	0.556	ug/l	90
9) 1,1,2-Trichlorotrifluoroet	3.72	101	1492655	582.819	ug/l	99
12) 1,1-Dichloroethene	3.70	96	10880	4.194	ug/l	90
27) cis-1,2-Dichloroethene	6.78	96	13553	4.166	ug/l	88
44) Trichloroethene	8.80	130	538438	170.287	ug/l	98
45) 1,2-Dichloropropane	9.08	63	2664	0.750	ug/l	93
49) 1,4-Dioxane	9.16	88	1640	29.023	ug/l #	78
64) Tetrachloroethene	10.60	164	208434	70.710	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092120\
Data File : VN063515.D
Acq On : 22 Sep 2020 4:06
Operator : JC/MD
Sample : L4082-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE134D3-20200915

Quant Time: Sep 22 09:33:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

