

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092120\
 Data File : VN063516.D
 Acq On : 22 Sep 2020 4:29
 Operator : JC/MD
 Sample : L4082-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE134D4-20200915

Quant Time: Sep 22 09:36:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	221581	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	414625	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	410358	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	176641	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	183456	48.64	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	97.28%	
35) Dibromofluoromethane	7.55	113	132911	46.30	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	92.60%	
50) Toluene-d8	10.06	98	527161	48.84	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	97.68%	
62) 4-Bromofluorobenzene	12.38	95	208884	52.79	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	105.58%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	3.72	101	724994	293.672	ug/l	99
12) 1,1-Dichloroethene	3.71	96	4610	1.844	ug/l #	58
27) cis-1,2-Dichloroethene	6.78	96	3788	1.208	ug/l	77
44) Trichloroethene	8.80	130	92555	30.263	ug/l	98
49) 1,4-Dioxane	9.17	88	382	6.989	ug/l	90
52) Toluene	10.12	92	3184	0.420	ug/l	100
64) Tetrachloroethene	10.60	164	36756	13.154	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092120\
Data File : VN063516.D
Acq On : 22 Sep 2020 4:29
Operator : JC/MD
Sample : L4082-11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE134D4-20200915

Quant Time: Sep 22 09:36:04 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

