

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092120\
 Data File : VN063519.D
 Acq On : 22 Sep 2020 5:41
 Operator : JC/MD
 Sample : L4082-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP01-20200915

Quant Time: Sep 22 09:40:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	219478	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	409696	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	404176	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	169504	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	180947	48.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.86%	
35) Dibromofluoromethane	7.55	113	130923	46.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.32%	
50) Toluene-d8	10.06	98	515556	48.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.68%	
62) 4-Bromofluorobenzene	12.38	95	201230	51.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.94%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.72	101	12101	4.949	ug/l	99
12) 1,1-Dichloroethene	3.70	96	8067	3.257	ug/l #	93
27) cis-1,2-Dichloroethene	6.78	96	5177	1.667	ug/l	86
30) Chloroform	7.33	83	4315	0.788	ug/l	97
38) Carbon Tetrachloride	7.73	117	5219	1.171	ug/l	93
44) Trichloroethene	8.80	130	487189	161.212	ug/l	98
55) 1,1,2-Trichloroethane	10.53	97	2762	0.928	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092120\
Data File : VN063519.D
Acq On : 22 Sep 2020 5:41
Operator : JC/MD
Sample : L4082-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP01-20200915

Quant Time: Sep 22 09:40:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

