

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092121\
 Data File : VN068580.D
 Acq On : 21 Sep 2021 14:20
 Operator : JC/MD
 Sample : M3781-10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB01-20210915

Quant Time: Sep 21 16:09:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	660713	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	1029688	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	941000	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	352452	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	341628	56.368	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	112.740%
35) Dibromofluoromethane	8.024	113	306423	49.341	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.680%
50) Toluene-d8	10.446	98	1203006	47.428	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.860%
62) 4-Bromofluorobenzene	12.734	95	353671	42.778	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	85.560%
Target Compounds						
					Qvalue	
89) n-Butylbenzene	13.949	91	3212	0.213	ug/l	94
95) Naphthalene	15.507	128	8060	4.034	ug/l	97
96) 1,2,3-Trichlorobenzene	15.705	180	2693	3.382	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092121\
Data File : VN068580.D
Acq On : 21 Sep 2021 14:20
Operator : JC/MD
Sample : M3781-10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB01-20210915

Quant Time: Sep 21 16:09:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

