

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092121\
 Data File : VN068581.D
 Acq On : 21 Sep 2021 14:46
 Operator : JC/MD
 Sample : M3781-09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB01-20210915

Quant Time: Sep 21 16:10:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	787809	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	1133137	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	980120	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	367964	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	404097	55.918	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 111.840%			
35) Dibromofluoromethane	8.024	113	359780	52.643	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 105.280%			
50) Toluene-d8	10.446	98	1362411	48.705	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 97.420%			
62) 4-Bromofluorobenzene	12.733	95	386819	42.542	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 85.080%			
Target Compounds						Qvalue
11) Tert butyl alcohol	5.369	59	14822	1.970	ug/l	99
16) Acetone	4.296	43	26518	3.942	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092121\
Data File : VN068581.D
Acq On : 21 Sep 2021 14:46
Operator : JC/MD
Sample : M3781-09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB01-20210915

Quant Time: Sep 21 16:10:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

