

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092121\
 Data File : VN068583.D
 Acq On : 21 Sep 2021 15:36
 Operator : JC/MD
 Sample : M3781-06DL 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW180D1-20210915DL

Quant Time: Sep 21 16:11:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	716021	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	1055802	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	922264	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	332260	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	360735	54.923	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.840%	
35) Dibromofluoromethane	8.027	113	325118	51.056	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 102.120%	
50) Toluene-d8	10.443	98	1235419	47.496	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 95.000%	
62) 4-Bromofluorobenzene	12.733	95	350781	41.515	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 83.040%	
Target Compounds						
44) Trichloroethene	9.220	130	134798	16.340	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092121\
Data File : VN068583.D
Acq On : 21 Sep 2021 15:36
Operator : JC/MD
Sample : M3781-06DL 5X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW180D1-20210915DL

Quant Time: Sep 21 16:11:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

