

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092121\
 Data File : VN068587.D
 Acq On : 21 Sep 2021 17:17
 Operator : JC/MD
 Sample : M3781-14DL 25X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE115D2-20210916DL

Quant Time: Sep 21 17:55:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	665450	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	979298	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	893703	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	348892	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	339324	55.589	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.180%
35) Dibromofluoromethane	8.027	113	303016	51.303	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.600%
50) Toluene-d8	10.443	98	1160084	48.040	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.080%
62) 4-Bromofluorobenzene	12.733	95	334538	42.569	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	85.140%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.213	101	10198	1.786	ug/l	# 85
12) 1,1-Dichloroethene	4.194	96	6035	1.154	ug/l	97
44) Trichloroethene	9.220	130	285072	37.256	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092121\
Data File : VN068587.D
Acq On : 21 Sep 2021 17:17
Operator : JC/MD
Sample : M3781-14DL 25X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE115D2-20210916DL

Quant Time: Sep 21 17:55:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

