

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092121\
 Data File : VN068588.D
 Acq On : 21 Sep 2021 17:42
 Operator : JC/MD
 Sample : M3781-16DL 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE134D3-20210916DL

Quant Time: Sep 21 18:16:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	586834	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	917889	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	813118	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	305938	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	322293	59.872	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 119.740%			
35) Dibromofluoromethane	8.024	113	283911	51.284	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 102.560%			
50) Toluene-d8	10.443	98	1099576	48.540	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 97.080%			
62) 4-Bromofluorobenzene	12.734	95	319622	43.311	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 86.620%			
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.218	101	114129	22.662	ug/l	96
44) Trichloroethene	9.220	130	67613	9.427	ug/l	91
64) Tetrachloroethene	10.977	164	14220	1.973	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092121\
Data File : VN068588.D
Acq On : 21 Sep 2021 17:42
Operator : JC/MD
Sample : M3781-16DL 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE134D3-20210916DL

Quant Time: Sep 21 18:16:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

