

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092121\
 Data File : VN068589.D
 Acq On : 21 Sep 2021 18:07
 Operator : JC/MD
 Sample : M3781-15
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE134D1-20210916

Quant Time: Sep 21 18:24:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	647709	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	987206	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	877828	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	334050	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	349228	58.778	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	117.560%
35) Dibromofluoromethane	8.024	113	309866	52.042	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.080%
50) Toluene-d8	10.446	98	1214562	49.755	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.520%
62) 4-Bromofluorobenzene	12.736	95	349826	44.002	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.000%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.220	130	30733	3.984	ug/l	98
64) Tetrachloroethene	10.979	164	3661	0.471	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092121\
Data File : VN068589.D
Acq On : 21 Sep 2021 18:07
Operator : JC/MD
Sample : M3781-15
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE134D1-20210916

Quant Time: Sep 21 18:24:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

