

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092121\
 Data File : VN068591.D
 Acq On : 21 Sep 2021 18:58
 Operator : JC/MD
 Sample : PB139153ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB139153ZHE#01

Quant Time: Sep 22 03:43:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	653029	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	1004679	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	1012043	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	371513	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	341257	56.969	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	113.940%
35) Dibromofluoromethane	8.024	113	304431	50.240	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.480%
50) Toluene-d8	10.446	98	1254283	50.436	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.880%
62) 4-Bromofluorobenzene	12.734	95	366369	45.161	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.320%
Target Compounds						
						Qvalue
16) Acetone	4.296	43	47191	19.737	ug/l	93
20) Methylene Chloride	5.103	84	28518	3.828	ug/l	96
43) Isopropyl Acetate	8.563	43	47093	4.189	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092121\
Data File : VN068591.D
Acq On : 21 Sep 2021 18:58
Operator : JC/MD
Sample : PB139153ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB139153ZHE#01

Quant Time: Sep 22 03:43:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

