

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092121\
 Data File : VN068592.D
 Acq On : 21 Sep 2021 19:23
 Operator : JC/MD
 Sample : PB139153ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB139153ZHE#02

Quant Time: Sep 22 03:43:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	775393	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	1029625	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	946358	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	372499	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	404638	56.890	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 113.780%			
35) Dibromofluoromethane	8.027	113	357512	57.571	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 115.140%			
50) Toluene-d8	10.443	98	1202033	47.395	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 94.800%			
62) 4-Bromofluorobenzene	12.733	95	411895	49.140	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 98.280%			
Target Compounds						Qvalue
16) Acetone	4.293	43	46213	14.555	ug/l	94
20) Methylene Chloride	5.103	84	29698	3.183	ug/l	96
43) Isopropyl Acetate	8.563	43	60276	5.232	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092121\
Data File : VN068592.D
Acq On : 21 Sep 2021 19:23
Operator : JC/MD
Sample : PB139153ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB139153ZHE#02

Quant Time: Sep 22 03:43:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

