

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092121\
 Data File : VN068593.D
 Acq On : 21 Sep 2021 19:48
 Operator : JC/MD
 Sample : PB139153ZHE#03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB139153ZHE#03

Quant Time: Sep 22 03:43:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	676623	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	1029780	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	983457	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	389610	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	347561	55.998	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 112.000%			
35) Dibromofluoromethane	8.027	113	315902	50.862	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.720%			
50) Toluene-d8	10.446	98	1259704	49.491	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 98.980%			
62) 4-Bromofluorobenzene	12.733	95	390148	46.758	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 93.520%			
Target Compounds					Qvalue	
16) Acetone	4.293	43	47718	19.025	ug/l	98
20) Methylene Chloride	5.111	84	30216	3.947	ug/l	97
43) Isopropyl Acetate	8.563	43	53857	4.674	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092121\
Data File : VN068593.D
Acq On : 21 Sep 2021 19:48
Operator : JC/MD
Sample : PB139153ZHE#03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB139153ZHE#03

Quant Time: Sep 22 03:43:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

