

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092121\
 Data File : VN068596.D
 Acq On : 21 Sep 2021 21:04
 Operator : JC/MD
 Sample : M3772-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PRE-CLASS-WASTE-1

Quant Time: Sep 22 03:44:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	884401	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	1308616	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	1181788	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	439805	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	473408	58.355	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 116.700%			
35) Dibromofluoromethane	8.027	113	401888	50.919	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.840%			
50) Toluene-d8	10.443	98	1557242	48.242	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 96.480%			
62) 4-Bromofluorobenzene	12.734	95	460412	43.718	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 87.440%			
Target Compounds					Qvalue	
16) Acetone	4.293	43	97247	35.160	ug/l	95
20) Methylene Chloride	5.106	84	42363	4.336	ug/l	94
27) cis-1,2-Dichloroethene	7.329	96	143209	15.371	ug/l	95
43) Isopropyl Acetate	8.568	43	77292	5.279	ug/l #	78
44) Trichloroethene	9.217	130	40437	3.955	ug/l	86
64) Tetrachloroethene	10.980	164	250530	23.919	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092121\
Data File : VN068596.D
Acq On : 21 Sep 2021 21:04
Operator : JC/MD
Sample : M3772-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PRE-CLASS-WASTE-1

Quant Time: Sep 22 03:44:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

