

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092122\
 Data File : VN074603.D
 Acq On : 21 Sep 2022 10:32
 Operator : JC\MD
 Sample : VN0921MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0921MBL01

Quant Time: Sep 22 06:56:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	312873	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	590703	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	478415	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	172707	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.363	65	272833	51.394	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.780%
35) Dibromofluoromethane	7.945	113	198905	50.690	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.380%
50) Toluene-d8	10.381	98	693125	45.675	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.340%
62) 4-Bromofluorobenzene	12.674	95	205836	39.603	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	79.200%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092122\
Data File : VN074603.D
Acq On : 21 Sep 2022 10:32
Operator : JC\MD
Sample : VN0921MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0921MBL01

Quant Time: Sep 22 06:56:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

