

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092122\
 Data File : VN074607.D
 Acq On : 21 Sep 2022 12:35
 Operator : JC\MD
 Sample : N4645-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TOLL-MW-08

Quant Time: Sep 22 06:57:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/22/2022
 Supervised By : Mahesh Dadoda 09/22/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	340968	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	672521	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	638354	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	254289	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	265631	45.914	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	91.820%		
35) Dibromofluoromethane	7.951	113	213424	47.773	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.540%		
50) Toluene-d8	10.380	98	781732	45.246	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	90.500%		
62) 4-Bromofluorobenzene	12.674	95	284115	48.014	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	96.020%		
Target Compounds						
19) Methyl tert-butyl Ether	5.510	73	19969m	1.055	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092122\
Data File : VN074607.D
Acq On : 21 Sep 2022 12:35
Operator : JC\MD
Sample : N4645-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TOLL-MW-08

Quant Time: Sep 22 06:57:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 09/22/2022
Supervised By :Mahesh Dadoda 09/22/2022

