

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092122\
 Data File : VN074608.D
 Acq On : 21 Sep 2022 13:00
 Operator : JC\MD
 Sample : N4645-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TOLL-MW-09

Quant Time: Sep 22 06:57:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	409893	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	765129	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	775697	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	348322	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	303826	43.686	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.380%
35) Dibromofluoromethane	7.951	113	251986	49.578	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.160%
50) Toluene-d8	10.380	98	919244	46.766	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.540%
62) 4-Bromofluorobenzene	12.674	95	346813	51.515	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.040%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.281	59	129516	57.167	ug/l	96
16) Acetone	4.234	43	10291	2.312	ug/l	96
19) Methyl tert-butyl Ether	5.534	73	90598	3.982	ug/l #	73
22) Diisopropyl ether	6.410	45	21770	0.841	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092122\
Data File : VN074608.D
Acq On : 21 Sep 2022 13:00
Operator : JC\MD
Sample : N4645-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TOLL-MW-09

Quant Time: Sep 22 06:57:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

